

INDIAN INSTITUTE OF TECHNOLOGY KANPUR

Centre for Environmental Science and Engineering

Enquiry No. IITK/CESE/2015-2016/1

Date 20/11/15

Sub: Quotation invited for test-tube rotator

CESE, IIT Kanpur invites quotations from companies and firms for test tube rotator with the following specification

1. **Sturdy construction and compact design**
2. **Holds all standard tube sizes (with adjustable clip discs)**
3. **Adjustable angles**
4. **Ambient operating range 5°C to 60°C**
5. **Timing range 1 – 99 min**
6. **RPM upto 40**
7. **Smooth adjustable speed control**
8. **1 Year warranty**

The sealed quotations addressed to Dr. Nalini Sankararamakrishnan, CESE should be sent to the following address latest by 28th Nov 2015. Please attach equipment brochure along with your quotation.

Dr. Nalini Sankararamakrishnan
Centre for Environmental Science and Engineering
Indian Institute of Technology Kanpur
Kanpur U.P. 208016,