



Enquiry No.: SSK/CHE/01/2016-17

Enquiry Date: December 05, 2016

Closing Date: December 12, 2016

Sealed quotation(s) in Indian Rupees or USD with all technical details so as to reach latest by 3:00 PM on December 12, 2016 are invited for the supply of following items.

Note: Price Bid and Technical bid of instrument should be provided separately with same date and also mention the enquiry number and instrument name on the sealed envelope carrying the quotation. Please provide all the mentioned parts together with name of the companies, specified.

High Pressure Reactor

Specification:-

Temperature up to 800 °C, ICONAL Tube I.D. 25 mm, heating zone 200 mm, total furnace length 400 mm. best quality heating elements proper ceramic fiber Insulation and separate control box having MCB Switch, indicator lights, PID profile programmer of 16 steps, with Cr/Al thermocouple inserted from top hole of Furnace.

Pre Heater

Pre heater just before the reactor to heat mixture upto 400°C with temperature controller

Thermal Mass Flow Controllers for He & H₂ Gases

(a) Gas – Helium

- Op. Temp.: 30-100 deg c
- Inlet Pressure 30 bar (g)
- Outlet Pressure 20 bar (g)
 - Flow-0 –4 SLPM
 - MOC-SS 316L
- End Connection-¼” OD compression

(b) Gas–hydrogen

- Op. Temp.: 30-100 deg c
- Inlet Pressure: 30 bar (g)
- Outlet Pressure : 20 bar (g)
 - Flow-0 –1 SLPM
 - MOC-SS 316L
- End Connection-¼” OD compression

Double Stage High Pressure Regulators (2 Nos.)

SS 316 Dual Stage Regulators with Two Dial Gauge upto 100 Bar

Micro Fluid Pump: (1 No)

Capable of Pumping Solvents like acids, bases, biomass Reactants like furfural, hmf, etc., at least in the range of (0.1 – 10 ml per min) at Room Temperature to 50 °C

Condenser with Water Chiller (1 No)

Condenser Connect to Chiller Machine, for Cooling Outlet from the Reactor.

Temperature Range: -5 °C to RT

Material: Glass

Pump Capacity: 12 lit/min

Cooling Capacity : 0.25KW

Tank Capacity 20 liters

Chiller having suitable specs will be accepted

Filters (2 Nos)

Micron Filters which can collect Solids (Range 0.100 - 0.250 Microns) from Reactor and avoid them going into back pressure Regulator.

One additional 5 micron Filter will be fixed before fine filter to increase life of Second Filter.

Liquid Gas Separator (1 No)

Liquid Gas Separator made up of Glass and Connected to GC

Back Pressure Regulator (2 Nos)

Medium to High Pressure Piston up to 100 Bar Pressure, compatible with various solvents (Acids, Bases, Biomass Reactants like furfural, hmf etc)

Pressure Gauge: (1 No)

Pressure Gauge (0-100 Bar) for measuring actual pressure inside the Reactor

High Pressure ON / OFF Valve (2 Nos)

Suitable High Pressure ON / OFF Valve to stop the Gas and Fluid Flow Completely

Non Return Valves (2 Nos)

Suitable High Pressure NRVs to prevent back flow of Gas and Fluid. Gas and Fluid Pipelines:

High Purity SS 316 1/8 inch tubes with fittings to sustain High Pressure of 100 Bars

The machine should be duly certified / authorized and the vendor should produce the certificate for the same.

Terms & Conditions:

- I. Price Bid and Technical bid should be provided separately with same date.
- II. Prices (FOB/ High Sea Sales) should include delivery up to nearest airport.
- III. Clearly state the CIF charges to IIT Kanpur and other taxes as applicable.
- IV. Warranty should at least be for 1-3 years after installation.
- V. Validity of quotation should be at least for 90 days.
- VI. The delivery time should be clearly mentioned. Shorter delivery time may be given a preference.
- VII. Technical specifications along with the extent of compliance should be in a separate envelope with proper labels on the envelopes.
- VIII. The delivery period should be specifically stated.

Kindly mention the enquiry number on the sealed envelope carrying the quotation and send the sealed quotation(s) to the following address:

Dr. Sri Sivakumar
Department of Chemical Engineering
Indian Institute of Technology Kanpur 208016
Kanpur, U.P., INDIA
Phone No. +91-512-259 7697/7895

